

Water Vapor 5/a-P

Order No. 67 28 531

Application Range

Use in Aerotest 5000, SimultanTest CO₂

Standard Measuring Range: 5 to 200 mg/m³

Test Volume: 50 L

Flow Rate: 2 L / min

Time for Measurement: approx. 25 min

Standard Deviation: ± 15 to 20 %

Color Change: yellow → reddish-brown

Ambient Operating Conditions

Temperature: 0 to 40 °C

Pressure: The tube may only be used for depressurized compressed air

Reaction Principle

$\text{H}_2\text{O} + \text{SeO}_2 + \text{H}_2\text{SO}_4 \rightarrow$ reddish-brown reaction product

Cross Sensitivity

Alcohols and unsaturated hydrocarbons of high concentrations may cause a diffused discoloration of the indicating layer.

Extension of the measuring range

The following evaluation applies for other volumes:

reading:	5	10	30	50	70	100	150	200	mg H ₂ O/m ³
----------	---	----	----	----	----	-----	-----	-----	------------------------------------

25 L vol.:	10	20	70	110	160	220	340	450	mg H ₂ O/m ³
------------	----	----	----	-----	-----	-----	-----	-----	------------------------------------

100 L vol.:	2	4	12	20	28	40	60	80	mg H ₂ O/m ³
-------------	---	---	----	----	----	----	----	----	------------------------------------

i.e. given a test volume of 25 L the scale reading of 50 mg H₂O/m³ corresponds to a measured value of 110 mg H₂O/m³

Relative Standard Deviation: ± 25 to 30 % (25 L)
± 20 to 25 % (100 L)

